

SUPERVISOR TEST



LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 07/06/2025

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:07
INTERNAL STANDARD	VERIFIED	12:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:08
BLANK TEST	0.00	12:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:09
BLANK TEST	0.00	12:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:10
BLANK TEST	0.00	12:11



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 07/13/2025

OPERATOR NAME:
MAKENZIE A HAYES
Permit #: 49532
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:26
INTERNAL STANDARD	VERIFIED	13:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:26
BLANK TEST	0.00	13:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:27
BLANK TEST	0.00	13:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:28
BLANK TEST	0.00	13:29



SUPERVISOR TEST


LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 07/21/2025

OPERATOR NAME:
MAKENZIE A HAYES
Permit #: 49532
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:22
INTERNAL STANDARD	VERIFIED	06:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:23
BLANK TEST	0.00	06:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:24
BLANK TEST	0.00	06:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:25
BLANK TEST	0.00	06:26



SOLUTION CHANGE

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 07/27/2025

OPERATOR NAME:
MAKENZIE A HAYES
Permit #: 49532
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:02
INTERNAL STANDARD	VERIFIED	19:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:03
BLANK TEST	0.00	19:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:04
BLANK TEST	0.00	19:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:05
BLANK TEST	0.00	19:06

