

SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 02/04/2019

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:58
INTERNAL STANDARD	VERIFIED	12:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:59
BLANK TEST	0.00	13:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:00
BLANK TEST	0.00	13:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:01
BLANK TEST	0.00	13:02



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 02/10/2019

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:28
INTERNAL STANDARD	VERIFIED	12:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:28
BLANK TEST	0.00	12:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:29
BLANK TEST	0.00	12:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:31
BLANK TEST	0.00	12:31



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 02/18/2019

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:25
INTERNAL STANDARD	VERIFIED	06:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:25
BLANK TEST	0.00	06:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:27
BLANK TEST	0.00	06:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:28
BLANK TEST	0.00	06:29



SUPERVISOR TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 02/24/2019

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:17
INTERNAL STANDARD	VERIFIED	10:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:17
BLANK TEST	0.00	10:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:18
BLANK TEST	0.00	10:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:20
BLANK TEST	0.00	10:20

