

SUPERVISOR TEST




LAKE PLACID VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 100290

Date: 09/05/2021

OPERATOR NAME:  
 CHAD P BLINN  
 Permit #: 33034  
 Operator Agency/Dept: LAKE PLACID VILL. PD  
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:04
INTERNAL STANDARD	VERIFIED	09:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:05
BLANK TEST	0.00	09:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:06
BLANK TEST	0.00	09:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:07
BLANK TEST	0.00	09:08



SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 09/13/2021

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:34
INTERNAL STANDARD	VERIFIED	10:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:34
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:35
BLANK TEST	0.00	10:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:36
BLANK TEST	0.00	10:37



SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 09/19/2021

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:16
INTERNAL STANDARD	VERIFIED	07:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:17
BLANK TEST	0.00	07:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:18
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:19
BLANK TEST	0.00	07:20



SOLUTION CHANGE

LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 09/27/2021

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:25
INTERNAL STANDARD	VERIFIED	08:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:26
BLANK TEST	0.00	08:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:27
BLANK TEST	0.00	08:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:28
BLANK TEST	0.00	08:29

