

SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 02/02/2025

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:17
INTERNAL STANDARD	VERIFIED	02:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:17
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:20
BLANK TEST	0.00	02:20



SUPERVISOR TEST




LAKE PLACID VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 100290

Date: 02/09/2025

OPERATOR NAME:  
 MAKENZIE A HAYES  
 Permit #: 49532  
 Operator Agency/Dept: LAKE PLACID VILL. PD  
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:01
INTERNAL STANDARD	VERIFIED	15:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:01
BLANK TEST	0.00	15:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:02
BLANK TEST	0.00	15:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:04
BLANK TEST	0.00	15:04



SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 02/16/2025

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:40
INTERNAL STANDARD	VERIFIED	04:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:40
BLANK TEST	0.00	04:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:42
BLANK TEST	0.00	04:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:43
BLANK TEST	0.00	04:43



SUPERVISOR TEST

LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 02/23/2025

OPERATOR NAME:  
MAKENZIE A HAYES  
Permit #: 49532  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:11
INTERNAL STANDARD	VERIFIED	08:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:12
BLANK TEST	0.00	08:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:13
BLANK TEST	0.00	08:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:14
BLANK TEST	0.00	08:15

