

SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 01/05/2025

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:29
INTERNAL STANDARD	VERIFIED	05:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:29
BLANK TEST	0.00	05:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:30
BLANK TEST	0.00	05:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:31
BLANK TEST	0.00	05:32



SUPERVISOR TEST


LAKE PLACID VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 100290

Date: 01/12/2025

OPERATOR NAME:
 MAKENZIE A HAYES
 Permit #: 49532
 Operator Agency/Dept: LAKE PLACID VILL. PD
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:35
INTERNAL STANDARD	VERIFIED	13:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:36
BLANK TEST	0.00	13:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:37
BLANK TEST	0.00	13:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:38
BLANK TEST	0.00	13:39



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 01/19/2025

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:14
INTERNAL STANDARD	VERIFIED	05:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:15
BLANK TEST	0.00	05:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:16
BLANK TEST	0.00	05:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:17
BLANK TEST	0.00	05:18



SOLUTION CHANGE

LAKE PLACID VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 100290

Date: 01/26/2025

OPERATOR NAME:
 MAKENZIE A HAYES
 Permit #: 49532
 Operator Agency/Dept: LAKE PLACID VILL. PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:07
INTERNAL STANDARD	VERIFIED	14:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:07
BLANK TEST	0.00	14:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:08
BLANK TEST	0.00	14:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:10
BLANK TEST	0.00	14:10

