



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NATL PATENT ANALYTICAL SYSTEM, DATAMASTER DMT breath test instrument, serial number 100290, performed on February 07, 2022 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 02/22/2022 11:05 AM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake Placid Village Police Department
Manufacturer: NATL PATENT ANALYTICAL SYSTEM
Model: DATAMASTER DMT
Serial Number: 100290
Date: February 07, 2022

Reference Solution Lot Number : 21200
Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.099 %	34.00 degrees centigrade
2	0.099 %	34.00 degrees centigrade
3	0.100 %	34.00 degrees centigrade
4	0.099 %	34.00 degrees centigrade

Reference Solution Lot Number : 21310
Reference Solution Standard : 0.08% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.079 %	34.01 degrees centigrade
2	0.080 %	34.01 degrees centigrade
3	0.080 %	34.01 degrees centigrade
4	0.079 %	34.01 degrees centigrade

Reference Solution Lot Number : 21130
Reference Solution Standard : 0.18% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.179 %	34.02 degrees centigrade
2	0.179 %	34.02 degrees centigrade
3	0.181 %	34.02 degrees centigrade
4	0.180 %	34.02 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NATL PATENT ANALYTICAL SYSTEM breath test instrument model DATAMASTER DMT, serial number 100290, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under ESRA
by Jonathan Mascolo
on 02/22/2022 10:53 AM

Reviewed by:



Digitally signed under ESRA
by Frank Sgarlata
on 02/22/2022 11:01 AM

Jonathan Mascolo
Highway Safety Equipment Technician
Breath Analysis Operator Permit #45339

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake Placid Village Police Department
Manufacturer: NATL PATENT ANALYTICAL SYSTEM
Model: DATAMASTER DMT
Serial Number: 100290
Date: February 07, 2022

- 1. General appearance and condition of this device -----: **OK**

- 2. Check controls and indicators -----: **OK**

- 3. Check breath flow system -----: **OK**

- 4. Check printer operation -----: **OK**

- 5. System cycles through all modes -----: **OK**



6. Record any repairs made to this instrument and list services provided : INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES PRINT AND REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE, RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION.

Performed by:

Handwritten signature of Jonathan Mascolo in black ink.

Digitally signed under ESRA
by Jonathan Mascolo
on 02/22/2022 10:53 AM

Jonathan Mascolo
Highway Safety Equipment Technician
Breath Analysis Operator Permit #45339

Reviewed by:

Handwritten signature of Frank Sgarlata in black ink.

Digitally signed under ESRA
by Frank Sgarlata
on 02/22/2022 11:01 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

TECHNICIAN TEST

LAKE PLACID VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 100290





Date: 02/07/2022

OPERATOR NAME:
 JONATHAN M MASCOLO
 Permit #: 45339
 Operator Agency/Dept: LAKE PLACID VILL. PD
 Reference Standard #: 21200

Nominal = 0.100 g/210L

BLANK TEST	0.000	11:11
INTERNAL STANDARD	VERIFIED	11:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.099	11:12
BLANK TEST	0.000	11:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.099	11:13
BLANK TEST	0.000	11:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.100	11:14
BLANK TEST	0.000	11:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.099	11:15
BLANK TEST	0.000	11:16

Average = 0.0992
 Std Dev = 0.0005



TECHNICIAN TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290


Date: 02/07/2022

OPERATOR NAME:
JONATHAN M MASCOLO
Permit #: 45339
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 21310

Nominal = 0.080 g/210L

BLANK TEST	0.000	11:04
INTERNAL STANDARD	VERIFIED	11:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.079	11:04
BLANK TEST	0.000	11:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.080	11:05
BLANK TEST	0.000	11:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.080	11:06
BLANK TEST	0.000	11:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.079	11:07
BLANK TEST	0.000	11:08

Average = 0.0795
Std Dev = 0.0006



TECHNICIAN TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290


Date: 02/07/2022

OPERATOR NAME:
JONATHAN M MASCOLO
Permit #: 45339
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 21130

Nominal = 0.180 g/210L

BLANK TEST	0.000	11:18
INTERNAL STANDARD	VERIFIED	11:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.179	11:19
BLANK TEST	0.000	11:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.179	11:20
BLANK TEST	0.000	11:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.181	11:21
BLANK TEST	0.000	11:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.180	11:22
BLANK TEST	0.000	11:23

Average = 0.1798
Std Dev = 0.0010




SERVICE AUTHORIZATION FORM
 Highway Safety Technology Unit
 80 South Swan Street
 Albany, NY 12210
 Ph: 518-485-7636 / 518-402-0689
 Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. PLEASE SIGN AND DATE FORM.

SECTION I - AGENCY INFORMATION

Name of Agency LAKE PLACID POLICE DEPARTMENT		Today's Date 2-07-2022
Address 2693 MAIN ST STE 104		City, State, ZIP LAKE PLACID, NY 12946
Contact Person PTL CHAD BLINN		Contact Telephone 518-523-3306
Contact Email blinn@lakeplacidpd.com or dobson@lakeplacidpd.com		Best Way to Reach You During Appointment Period NA

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT REPCO SIM Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH A/S FST Other:		
Model Number Datamaster DMT		Instrument Serial Number 100290
Antenna Serial Number (if applicable) Antenna 1 -- / Antenna 2 --		Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair
Brief Description of Malfunction 		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: <input checked="" type="checkbox"/> DMT <input checked="" type="checkbox"/> SIM <input checked="" type="checkbox"/> Keyboard Other:	Radar: <input checked="" type="checkbox"/> breath. tubes <input checked="" type="checkbox"/> Sim tubes	Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.		
Signature 		Date 2-07-2022

DCJS USE ONLY

Inspected instrument, checked Setup Voltages Print & Remote operation. Performed Calibration/Adjustment Procedure, Ran Subject, technician & Diagnostic tests, Verified Calibration <div align="right">* @ 2-7-2022</div>
--