




LAKE PLACID VILLAGE
POLICE
DATA # 100290

Date 025

OPERATOR:
MAKENZYES
Permit 32
Operator/Dept: LAKE PLACID VILL. PD
Reference # 25040

Nominal 0 g/210L

BLANK TEST	0.00	06:40
INTERNAL STANDARD	VERIFIED	06:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:40
BLANK TEST	0.00	06:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:42
BLANK TEST	0.00	06:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:43
BLANK TEST	0.00	06:43






LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 05/12/2025

OPERATOR NAME:
MAKENZIE A HAYES
Permit #: 49532
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:19
INTERNAL STANDARD	VERIFIED	18:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:19
BLANK TEST	0.00	18:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:20
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:21
BLANK TEST	0.00	18:22






LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 05/18/2025

OPERATOR NAME:
MAKENZIE A HAYES
Permit #: 49532
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:46
INTERNAL STANDARD	VERIFIED	06:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:46
BLANK TEST	0.00	06:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:48
BLANK TEST	0.00	06:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:49
BLANK TEST	0.00	06:50



SUPERVISOR TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 05/25/2025

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:23
INTERNAL STANDARD	VERIFIED	03:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:23
BLANK TEST	0.00	03:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:24
BLANK TEST	0.00	03:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:25
BLANK TEST	0.00	03:26

