

SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 03/06/2022

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:04
INTERNAL STANDARD	VERIFIED	16:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:05
BLANK TEST	0.00	16:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:06
BLANK TEST	0.00	16:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:07
BLANK TEST	0.00	16:08



SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 03/14/2022

OPERATOR NAME:  
DORAN H SORRELL  
Permit #: 47507  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:10
INTERNAL STANDARD	VERIFIED	18:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:10
BLANK TEST	0.00	18:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:12
BLANK TEST	0.00	18:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:13
BLANK TEST	0.00	18:13



SUPERVISOR TEST



LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 03/20/2022

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:27
INTERNAL STANDARD	VERIFIED	16:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:27
BLANK TEST	0.00	16:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:29
BLANK TEST	0.00	16:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:30
BLANK TEST	0.00	16:30



SOLUTION CHANGE

LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 03/28/2022

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:57
INTERNAL STANDARD	VERIFIED	07:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:57
BLANK TEST	0.00	07:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:58
BLANK TEST	0.00	07:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:00
BLANK TEST	0.00	08:00

