



**Division of Criminal  
Justice Services**

---

**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS  
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, **Steven J. Carluccio**, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of **GUTH LABORATORIES, 12V500 SIMULATOR**, serial number **MP4814**, performed on **February 07, 2022** by Office of Public Safety employee **Jonathan Mascolo**, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA  
by Steven J. Carluccio  
on 02/22/2022 03:23 PM

---

**Steven J. Carluccio**  
**Highway Safety Equipment Technical Supervisor**  
**Office of Public Safety**

**SIMULATOR  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency: Lake Placid Village Police Department**  
**Manufacturer: GUTH LABORATORIES**  
**Model: 12V500**  
**Serial Number: MP4814**  
**Date: February 07, 2022**

<b>Breath Simulator Temperature</b>
34.03 degrees centigrade



Division of Criminal  
Justice Services

**I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of GUTH LABORATORIES simulator model 12V500, serial number MP4814, and have determined that it is accurate.**

**I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified simulator was performed, or within a reasonable time thereafter.**

A handwritten signature in black ink that reads "Jonathan Mascolo".

Digitally signed under ESRA  
by Jonathan Mascolo  
on 02/22/2022 12:42 PM

---

**Jonathan Mascolo**  
**Highway Safety Equipment Technician**  
**Office of Public Safety**

**SIMULATOR  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency:** Lake Placid Village Police Department  
**Manufacturer:** GUTH LABORATORIES  
**Model:** 12V500  
**Serial Number:** MP4814  
**Date:** February 07, 2022

1. General appearance and condition of this device -----: **OK**

2. Check propeller rotation -----: **OK**

3. Check for air leaks -----: **OK**

4. Check for proper temperature -----: **OK**

5. Record any repairs made to this instrument and list services provided : **VERIFIED  
CALIBRATION**



*Jonathan Mascolo*

Digitally signed under ESRA  
by Jonathan Mascolo  
on 02/22/2022 12:42 PM

---

**Jonathan Mascolo  
Highway Safety Equipment Technician  
Office of Public Safety**

625 East Bunker Court  
 Vernon Hills, Illinois 60061  
 PH: 866-466-6225  
 Fax: 847-327-2993  
 www.innocalsolutions.com

## NIST Traceable Calibration Report



Reference Number: 1405165  
 PO Number: SCALUCCIO061421

New York State Division Criminal Justice Services  
 80 South Swan Street  
 Albany, NY 12210 United States

**Manufacturer:** Ertco  
**Model Number:** 4400  
**Description:** Temperature, Dig thermometer w/probe  
**Asset Number:** 304011  
**Serial Number:** 304011  
**Procedure:** DS Eutechnics 4400-EP642

**Calibration Date:** 06/17/2021  
**Calibration Due Date:** 06/17/2022  
**Condition As Found:** In Tolerance  
**Condition As Left:** In Tolerance, No adjustment

**Remarks:**

NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.

### Standards Utilized

Asset No.	Manufacturer	Model No.	Description	Cal. Date	Due Date
CP144317	Fluke Corporation	1595A	Thermometer, Readout, Precision Thermometer	01/28/2020	02/28/2022
CP32032	Hart Scientific	5628	Temperature, Probe	10/19/2020	04/30/2022
CP32034	Hart Scientific	5628	Temperature, Probe	01/22/2020	10/31/2021
CP32042	Hart Scientific	5680	Temperature, Probe, SPRT	03/31/2021	04/30/2022
CP53595	Hart Scientific	5628	Temperature, iProbe	01/28/2020	02/28/2022

### Calibration Data

FUNCTION TESTED	Nominal Value	As Found	Out of Tol	As Left	Out of Tol	CALIBRATION TOLERANCE
Temperature Accuracy	-15.000 °C	-14.97		Same		-15.050 to -14.950 °C [EMU 0.013 °C][TUR 3.8:1]
	0.000 °C	0.01		Same		-0.030 to 0.030 °C [EMU 0.012 °C][TUR 2.5:1]
	75.000 °C	75.01		Same		74.970 to 75.030 °C [EMU 0.015 °C][TUR 2.0:1]
	125.000 °C	124.99		Same		124.950 to 125.050 °C [EMU 0.015 °C][TUR 3.4:1]

Temperature: 21° C  
 Humidity: 49% RH  
 Rpt. No.: 1655081

Calibration Performed By:				Quality Reviewer:	
Name	ID #	Title	Phone	Name	Date
Sittinger, Jennifer	330	Metrologist	847-327-5317	Ziegler, Jeff	06/18/2021

This report may not be reproduced, stored in a retrieval system, or transmitted in any form or by any means, electronic, mechanical, photocopying, recording, or by any information storage and retrieval system, without the prior written permission of the manufacturer. The results shown in this report have been verified by the manufacturer. Measurements reported herein are traceable to NIST via NIST's standards maintained by NIST and were performed in accordance with NIST's NIST-17025 (CAL) and NIST-17026 (MPE) requirements. The accuracy of the results is based on the accuracy of the standards used and the accuracy of the measurement process. The accuracy of the results is based on the accuracy of the standards used and the accuracy of the measurement process. The accuracy of the results is based on the accuracy of the standards used and the accuracy of the measurement process.



**SERVICE AUTHORIZATION FORM**  
**Highway Safety Technology Unit**  
**80 South Swan Street**  
**Albany, NY 12210**  
**Ph: 518-485-7636 / 518-402-0689**  
**Fax: 518-457-6869**

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

**SECTION I - AGENCY INFORMATION**

Name of Agency <b>LAKE PLACID POLICE DEPARTMENT</b>		Today's Date <b>2-07-2022</b>
Address <b>2693 MAIN ST STE 104</b>		City, State, ZIP <b>LAKE PLACID, NY 12946</b>
Contact Person <b>PTL CHAD BLINN</b>		Contact Telephone <b>518-523-3306</b>
Contact Email <b>blinn@lakeplacidpd.com or dobson@lakeplacidpd.com</b>		Best Way to Reach You During Appointment Period <b>NA</b>

**SECTION II - INSTRUMENT INFORMATION**

Instrument Manufacturer (Please Circle) NPAS DMT   REPCO SIM   Applied Concepts(Stalker)   Decatur(Genesis)   Kustom(Eagle/Falcon)   MPH   A/S FST   Other:			
Model Number <b>Model 12V500</b>		Instrument Serial Number <b>MP4814</b>	
Antenna Serial Number (if applicable) Antenna 1 --   /   Antenna 2 --		Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction			
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)			
DMT(SIM) <b>(SIM)</b> <b>(Tubes)</b> DMT   Other: <b>(1) Power cable (1) Communication</b> Keyboard		Radar: Control Unit   Antenna(s)   Antenna Cable(s) Power Cord   Forks   Other:	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. <b>PLEASE SIGN AND DATE.</b>			
Signature <i>Chad Blinn</i>		Date <b>2-07-2022</b>	

**DCJS USE ONLY**

<u>Digital Thermo</u>	<u>DMT Export</u>	Air leak check - ok
34.03	34.01	Propeller - ok
34.03	34.01	Thermometer SN-304011
34.03	34.01	Exp: 6-17-2022

\*@ 2-7-2022