

SUPERVISOR TEST




-----  
LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290  
-----

Date: 07/08/2019

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:59
INTERNAL STANDARD	VERIFIED	06:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:59
BLANK TEST	0.00	07:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:01
BLANK TEST	0.00	07:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:02
BLANK TEST	0.00	07:03



SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 07/14/2019

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:36
INTERNAL STANDARD	VERIFIED	08:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:37
BLANK TEST	0.00	08:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:38
BLANK TEST	0.00	08:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:39
BLANK TEST	0.00	08:40



SUPERVISOR TEST




-----  
 LAKE PLACID VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 100290  
 -----

Date: 07/21/2019

OPERATOR NAME:  
 DORAN H SORRELL  
 Permit #: 47507  
 Operator Agency/Dept: LAKE PLACID VILL. PD  
 Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:01
INTERNAL STANDARD	VERIFIED	20:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:02
BLANK TEST	0.00	20:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:03
BLANK TEST	0.00	20:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:04
BLANK TEST	0.00	20:05



SOLUTION CHANGE

LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 07/29/2019

OPERATOR NAME:  
DORAN H SORRELL  
Permit #: 47507  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:11
INTERNAL STANDARD	VERIFIED	01:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:11
BLANK TEST	0.00	01:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:14

