

SUPERVISOR TEST




LAKE PLACID VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 100290

Date: 01/07/2019

OPERATOR NAME:
 CHAD P BLINN
 Permit #: 33034
 Operator Agency/Dept: LAKE PLACID VILL. PD
 Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:30
INTERNAL STANDARD	VERIFIED	08:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:30
BLANK TEST	0.00	08:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:32
BLANK TEST	0.00	08:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:33
BLANK TEST	0.00	08:34



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 01/13/2019

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:28
INTERNAL STANDARD	VERIFIED	05:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:28
BLANK TEST	0.00	05:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:29
BLANK TEST	0.00	05:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:30
BLANK TEST	0.00	05:31



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 01/21/2019

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:23
INTERNAL STANDARD	VERIFIED	17:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:23
BLANK TEST	0.00	17:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:25
BLANK TEST	0.00	17:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:26
BLANK TEST	0.00	17:27



SOLUTION CHANGE

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 01/27/2019

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:10
INTERNAL STANDARD	VERIFIED	17:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:10
BLANK TEST	0.00	17:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:12
BLANK TEST	0.00	17:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:13
BLANK TEST	0.00	17:14

