

SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 11/08/2021

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:16
INTERNAL STANDARD	VERIFIED	08:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:18
BLANK TEST	0.00	08:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:19
BLANK TEST	0.00	08:19



SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 11/14/2021

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:35
INTERNAL STANDARD	VERIFIED	16:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:36
BLANK TEST	0.00	16:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:37
BLANK TEST	0.00	16:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:38
BLANK TEST	0.00	16:39



SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 11/22/2021

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:49
INTERNAL STANDARD	VERIFIED	11:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:49
BLANK TEST	0.00	11:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:50
BLANK TEST	0.00	11:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:52
BLANK TEST	0.00	11:52



SOLUTION CHANGE

LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 11/28/2021

OPERATOR NAME:  
CHAD P BLINN  
Permit #: 33034  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:26
INTERNAL STANDARD	VERIFIED	16:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:26
BLANK TEST	0.00	16:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:27
BLANK TEST	0.00	16:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:29
BLANK TEST	0.00	16:29

