



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 100290, performed on February 19, 2021 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under
ESRA by Joshua Vinehout
on 3/1/2021 9:11 AM

Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake Placid Village Police Department

Manufacturer: NAT'L PATENT ANALYTICAL SYS Model: DATAMASTER DMT

Date: February 19, 2021 Serial Number: 100290

Reference Solution Lot Number: 20361

Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.097 %	34.00 degrees centigrade
2	0.098 %	34.00 degrees centigrade
3	0.098 %	34.00 degrees centigrade
4	0.097 %	34.00 degrees centigrade

Reference Solution Lot Number: 20470

Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.079 %	34.00 degrees centigrade
2	0.080 %	34.00 degrees centigrade
3	0.080 %	34.00 degrees centigrade
4	0.080 %	34.00 degrees centigrade

Reference Solution Lot Number: 20290


Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.178 %	34.00 degrees centigrade
2	0.177 %	34.00 degrees centigrade
3	0.178 %	34.00 degrees centigrade
4	0.178 %	34.00 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 100290, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under
ESRA by Jonathan Mascolo
on 2/19/2021 1:56 PM

Reviewed by:



Digitally signed under
ESRA by Frank Sgarlata
on 2/23/2021 10:33 AM

Jonathan Mascolo
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 45339

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake Placid Village Police Department
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 100290
Date: February 19, 2021

**1. General Appearance and condition of this device -----: INTERFERENCE
ERROR**

2. Check controls and indicators -----: OK

3. Check breath flow system -----: OK

4. Check printer operation -----: OK



5. System cycles through all modes -----: OK

**6. Record any repairs made to this instrument and list services provided: INSPECTED
INSTRUMENT, CLEANED FILTERS,CHECKED SETUP VOLTAGES PRINT AND
REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE,
RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION.**

Performed by:

Handwritten signature of Jonathan Mascolo in black ink.

Digitally signed under
ESRA by Jonathan Mascolo
on 2/19/2021 1:56 PM

Jonathan Mascolo
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 45339

Reviewed by:

Handwritten signature of Frank Sgarlata in black ink.

Digitally signed under
ESRA by Frank Sgarlata
on 2/23/2021 10:33 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

TECHNICIAN TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290





Date: 02/19/2021

OPERATOR NAME:
JONATHAN M MASCOLO
Permit #: 45339
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20361

Nominal = 0.100 g/210L

BLANK TEST	0.000	08:40
INTERNAL STANDARD	VERIFIED	08:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.097	08:41
BLANK TEST	0.000	08:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.098	08:42
BLANK TEST	0.000	08:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.098	08:43
BLANK TEST	0.000	08:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.097	08:44
BLANK TEST	0.000	08:45

Average = 0.0975
Std Dev = 0.0006



TECHNICIAN TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290





Date: 02/19/2021

OPERATOR NAME:
JONATHAN M MASCOLO
Permit #: 45339
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20470

Nominal = 0.080 g/210L

BLANK TEST	0.000	12:15
INTERNAL STANDARD	VERIFIED	12:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.079	12:15
BLANK TEST	0.000	12:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.080	12:16
BLANK TEST	0.000	12:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.080	12:18
BLANK TEST	0.000	12:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.080	12:19
BLANK TEST	0.000	12:20

Average = 0.0798
Std Dev = 0.0005



TECHNICIAN TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290





Date: 02/19/2021

OPERATOR NAME:
JONATHAN M MASCOLO
Permit #: 45339
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20290

Nominal = 0.180 g/210L

BLANK TEST	0.000	08:54
INTERNAL STANDARD	VERIFIED	08:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.178	08:55
BLANK TEST	0.000	08:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.177	08:56
BLANK TEST	0.000	08:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.178	08:57
BLANK TEST	0.000	08:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.178	08:58
BLANK TEST	0.000	08:59

Average = 0.1778
Std Dev = 0.0005



SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

2/19/21*

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. PLEASE SIGN AND DATE FORM.

SECTION I - AGENCY INFORMATION

Name of Agency LAKE PLACID POLICE DEPARTMENT		Today's Date 2-10-2021
Address 2693 MAIN ST		City, State, ZIP LAKE PLACID, NY 12946
Contact Person CHARLES LANGLOIS DELIVERING-CHAD BLINN @ STATION		Contact Telephone 518-523-3306
Contact Email blinn@lakeplacidpd.com		Best Way to Reach You During Appointment Period 518-523-3306

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT REPCO SIM Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH A/SFST Other.			
Model Number DMT*		Instrument Serial Number 100090*	
Antenna Serial Number (if applicable) Antenna 1 -- / Antenna 2 --		Type of maintenance (check appropriate) <input type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction Interference error*			
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)			
DMT/SIM: <input checked="" type="checkbox"/> DMT 100270 <input checked="" type="checkbox"/> SIM memory <input checked="" type="checkbox"/> Tubes <input checked="" type="checkbox"/> Keyboard Other: Power Cords		Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.			
Signature Pat Chad P. R. S.		Date 2-10-2021	

DCJS USE ONLY

<p>Inspected instrument, cleaned filters, checked setup voltages Print and Remote operation, Performed Calibration/Adjustment Procedure, Ran subject, technician & Diagnostic tests, Verified Calibration * @ 2/19/21</p>
