



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 100290, performed on February 04, 2020 by Office of Public Safety employee Joseph Iavarone, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance/calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under ESRA
by Joshua Vinehout
on 2/6/2020 12:18 PM

Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake Placid Village Police Department
Manufacturer: NAT'L PATENT ANALYTICAL SYS Model: DATAMASTER DMT
Serial Number: 100290
Date: February 04, 2020

Reference Solution Lot Number: 19350
Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.097 %	34.01 degrees centigrade
2	0.100 %	34.01 degrees centigrade
3	0.101 %	34.01 degrees centigrade
4	0.099 %	34.01 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 100290, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.


Performed by:



Digitally signed under
ESRA by Joseph Iavarone
on 2/4/2020 2:20 PM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 44106

Reviewed by:



Digitally signed under
ESRA by Frank Sgarlata
on 2/6/2020 11:41 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake Placid Village Police Department
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 100290
Date: February 04, 2020

1. General Appearance and condition of this device -----: N/A

2. Check controls and indicators -----: OK


3. Check breath flow system -----: OK

4. Check printer operation -----: N/A

5. System cycles through all modes -----: OK

6. Record any repairs made to this instrument and list services provided: **CHECKED REMOTE OPERATION, RAN DIAGNOSTIC AND TECHNICIAN TESTS, VERIFIED CALIBRATION.**


Performed by:



Digitally signed under
ESRA by Joseph Iavarone
on 2/4/2020 2:20 PM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 44106

Reviewed by:



Digitally signed under
ESRA by Frank Sgarlata
on 2/6/2020 11:41 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 02/04/2020

Nominal = 0.100 g/210L

BLANK TEST	0.000	13:20
INTERNAL STANDARD	VERIFIED	13:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.097	13:21
BLANK TEST	0.000	13:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.100	13:22
BLANK TEST	0.000	13:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.101	13:23
BLANK TEST	0.000	13:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.099	13:24
BLANK TEST	0.000	13:25

Average = 0.0992
Std Dev = 0.0017

Control Test Record

Serial Number: 100290

Test Date: 02-03-2020 16:02:47

Location: 324

LAKE PLACID POLICE DEPT

Supervisor Information

First Name: Chad P

Last Name: Blinn

Permit: 00104

Agency: LAKE PLACID VILL PD

Simulator Information

Lot: 19250

Nominal: 0.100

Test Results


Alcohol 1: 0.08 34.01

Alcohol 2: 0.10 34.01

Alcohol 3: 0.10 34.01

Solution Change:

Plot: 1 of 2



Status

Status Code:

Date Time:

Message:



Division of Criminal
Justice Services

**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 100290, performed on August 10, 2020 by Office of Public Safety employee Martin Schilling, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua J. Vinehout".

Digitally signed under
ESRA by Joshua Vinehout
on 8/17/2020 9:32 AM

Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

BREATH TEST INSTRUMENT RECORD OF INSPECTION / MAINTENANCE / CALIBRATION

Name of Submitting Agency: Lake Placid Village Police Department
 Manufacturer: NAT'L PATENT ANALYTICAL SYS Model: DATAMASTER DMT
 Date: August 10, 2020 Serial Number: 100290

Reference Solution Lot Number: 19281
 Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.100 %	34.01 degrees centigrade
2	0.102 %	34.01 degrees centigrade
3	0.101 %	34.01 degrees centigrade
4	0.101 %	34.01 degrees centigrade

Reference Solution Lot Number: 20020
 Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.080 %	34.01 degrees centigrade
2	0.082 %	34.01 degrees centigrade
3	0.080 %	34.01 degrees centigrade
4	0.080 %	34.01 degrees centigrade


Reference Solution Lot Number: 20290
 Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.182 %	34.01 degrees centigrade
2	0.181 %	34.01 degrees centigrade
3	0.182 %	34.01 degrees centigrade
4	0.182 %	34.01 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 100290, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.


Performed by:



Digitally signed under
 ESRA by Martin Schilling
 on 8/10/2020 12:19 PM

Martin Schilling
 Highway Safety Equipment Technician
 Breath Analysis Operator Permit # 45926

Reviewed by:



Digitally signed under
 ESRA by Frank Sgarlata
 on 8/14/2020 9:30 AM

Frank Sgarlata
 Breath Analysis Technical Supervisor


**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake Placid Village Police Department
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 100290
Date: August 10, 2020

- 1. General Appearance and condition of this device -----: OK
- 2. Check controls and indicators -----: OK
- 3. Check breath flow system -----: OK
- 4. Check printer operation -----: OK
- 5. System cycles through all modes -----: OK

6. Record any repairs made to this instrument and list services provided: **INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES PRINT AND REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE, RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION**

Performed by:



Digitally signed under
ESRA by Martin Schilling
on 8/10/2020 12:19 PM

Martin Schilling
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 45926

Reviewed by:



Digitally signed under
ESRA by Frank Sgarlata
on 8/14/2020 9:30 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

TECHNICIAN TEST

LAKE PLACID VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 100290





Date: 08/10/2020

OPERATOR NAME:
 MARTIN J SCHILLING
 Permit #: 45926
 Operator Agency/Dept: LAKE PLACID VILL. PD
 Reference Standard #: 19281

Nominal = 0.100 g/210L

BLANK TEST	0.000	10:56
INTERNAL STANDARD	VERIFIED	10:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.100	10:56
BLANK TEST	0.000	10:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.102	10:57
BLANK TEST	0.000	10:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.101	10:58
BLANK TEST	0.000	10:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.101	11:00
BLANK TEST	0.000	11:00

Average = 0.1010
 Std Dev = 0.0008



TECHNICIAN TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290





Date: 08/10/2020

OPERATOR NAME:
MARTIN J SCHILLING
Permit #: 45926
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20020

Nominal = 0.080 g/210L

BLANK TEST	0.000	11:04
INTERNAL STANDARD	VERIFIED	11:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.080	11:05
BLANK TEST	0.000	11:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.082	11:06
BLANK TEST	0.000	11:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.080	11:07
BLANK TEST	0.000	11:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.080	11:08
BLANK TEST	0.000	11:09

Average = 0.0805
Std Dev = 0.0010



TECHNICIAN TEST

LAKE PLACID VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 100290





Date: 08/10/2020

OPERATOR NAME:
 MARTIN J SCHILLING
 Permit #: 45926
 Operator Agency/Dept: LAKE PLACID VILL. PD
 Reference Standard #: 20290

Nominal = 0.180 g/210L

BLANK TEST	0.000	11:11
INTERNAL STANDARD	VERIFIED	11:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.182	11:12
BLANK TEST	0.000	11:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.181	11:13
BLANK TEST	0.000	11:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.182	11:14
BLANK TEST	0.000	11:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.182	11:16
BLANK TEST	0.000	11:16

Average = 0.1818
 Std Dev = 0.0005



New York State Division of Criminal Justice Services
SERVICE AUTHORIZATION FORM


MPJ
 8/10/2020

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE.

SECTION I - AGENCY INFORMATION

Name of Agency LAKE PLACID POLICE DEPARTMENT		Today's Date 8/10/2020
Address 2693 MAIN ST		City, State, ZIP LAKE PLACID, NY 12946
Contact Person PTL BLINN		Contact Telephone (518) 523-3306
Contact Telephone (518) 637-2572	Contact Facsimile (518) 523-9601	Contact Email blinn@lakeplacidpd.com

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer DATAMASTER		Model Number DMT
Instrument Serial Number 100290	Antenna Serial Number (if applicable)	
Date of Malfunction	Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units. MAIN UNIT Key Bel Simulator - 53007 - MP4814 CORDS		
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein.		
Signature 		Date 8/10/2020

DCJS USE ONLY

I inspected instrument. Checked set up voltages print and remote operation. Performed calibration/adjustment procedure. Per subject, technician and diagnostic tests. Verified calibration.