

SOLUTION CHANGE




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 11/29/2020

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:39
INTERNAL STANDARD	VERIFIED	16:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:40
BLANK TEST	0.00	16:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:41
BLANK TEST	0.00	16:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:42
BLANK TEST	0.00	16:43



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 12/07/2020

OPERATOR NAME:
DORAN H SORRELL
Permit #: 47507
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:03
INTERNAL STANDARD	VERIFIED	06:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:04
BLANK TEST	0.00	06:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:05
BLANK TEST	0.00	06:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:06
BLANK TEST	0.00	06:07



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 12/13/2020

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:39
INTERNAL STANDARD	VERIFIED	10:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:39
BLANK TEST	0.00	10:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:40
BLANK TEST	0.00	10:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:41
BLANK TEST	0.00	10:42



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 12/21/2020

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:44
INTERNAL STANDARD	VERIFIED	15:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:44
BLANK TEST	0.00	15:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:46
BLANK TEST	0.00	15:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:47
BLANK TEST	0.00	15:48



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 12/26/2020

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:50
INTERNAL STANDARD	VERIFIED	06:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:50
BLANK TEST	0.00	06:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:51
BLANK TEST	0.00	06:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:52
BLANK TEST	0.00	06:53



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 12/27/2020

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:08
INTERNAL STANDARD	VERIFIED	09:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:08
BLANK TEST	0.00	09:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:09
BLANK TEST	0.00	09:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:10
BLANK TEST	0.00	09:11



SUPERVISOR TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290




Date: 12/27/2020

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:32
INTERNAL STANDARD	VERIFIED	09:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:32
BLANK TEST	0.00	09:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:33
BLANK TEST	0.00	09:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	INTERFERENCE	09:34

Interference Detected



SUPERVISOR TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290



Date: 12/27/2020

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:46
INTERNAL STANDARD	VERIFIED	16:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:47
BLANK TEST	0.00	16:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	INTERFERENCE	16:48

Interference Detected



SUPERVISOR TEST

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 12/27/2020

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:04
INTERNAL STANDARD	VERIFIED	17:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:04
BLANK TEST	0.00	17:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:05
BLANK TEST	0.00	17:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:07
BLANK TEST	0.00	17:07

