

SUPERVISOR TEST



LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 03/06/2023

OPERATOR NAME:  
IAN O GORDON  
Permit #: 50784  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:57
INTERNAL STANDARD	VERIFIED	17:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:57
BLANK TEST	0.00	17:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:58
BLANK TEST	0.00	17:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:59
BLANK TEST	0.00	18:00



SUPERVISOR TEST




LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290

Date: 03/12/2023

OPERATOR NAME:  
IAN O GORDON  
Permit #: 50784  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:51
INTERNAL STANDARD	VERIFIED	05:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:51
BLANK TEST	0.00	05:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:53
BLANK TEST	0.00	05:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:54
BLANK TEST	0.00	05:54



SUPERVISOR TEST




-----  
LAKE PLACID VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 100290  
-----

Date: 03/20/2023

OPERATOR NAME:  
IAN O GORDON  
Permit #: 50784  
Operator Agency/Dept: LAKE PLACID VILL. PD  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:06
INTERNAL STANDARD	VERIFIED	18:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:07
BLANK TEST	0.00	18:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:08
BLANK TEST	0.00	18:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:09
BLANK TEST	0.00	18:10



SOLUTION CHANGE

-----  
 LAKE PLACID VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 100290  
 -----

Date: 03/26/2023

OPERATOR NAME:  
 IAN O GORDON  
 Permit #: 50784  
 Operator Agency/Dept: LAKE PLACID VILL. PD  
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:19
INTERNAL STANDARD	VERIFIED	18:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:19
BLANK TEST	0.00	18:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:20
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:21
BLANK TEST	0.00	18:22

